

Fresenius Medical Care North America Service Bulletin


Equipment: 2008K@Home Hemodialysis Machines
 Bulletin: 23-HOM-001 Rev A
 Subject: Repair Kits for *bi*bag® Connector

1.0 PURPOSE

To inform all Field Service and Technical Support personnel as well as Fresenius Medical Care-Trained customers of new kits to repair the *bi*bag® connector on *bi*bag® equipped 2008K@Home™ Hemodialysis Machines.

2.0 DESCRIPTION

Kits have been created to repair the *bi*bag® connector so it does not need to be completely replaced if a component breaks. The image and table below show the location, part number and description for each kit. Each kit includes instructions to complete the repair.



Part Number	Description
191349	<i>bi</i> bag Hall Switch Kit
191351	<i>bi</i> bag Locking Lever Kit
191352	<i>bi</i> bag Nozzle Housing Kit
191353	<i>bi</i> bag Flap Base Kit
191354	<i>bi</i> bag Sealing Bolt Kit
191355	<i>bi</i> bag Hinge Kit

© 2023 Fresenius Medical Care. All Rights Reserved.

If there are any questions regarding this bulletin, contact Fresenius Medical Care Technical Support at 800-227-2572.